

Experiment(s):


1. Determination of resistivity of a wire

(For more details, procedure & manual visit: www.kamaljeeth.net)

Reference : Lab Experiments Journal vol-14, No.2, Page-129

Experiment setup consists:

- Teak wood meter bridge
- Power supply
- Ammeter
- Voltmeter
- Rheostat
- Patch cords



(a)



(b)

**Specifications:****a) Teak wood meter bridge**

Material: Teak wood

Length: 1 m

Connectors: Brass with zinc coating

Test sample wire: Nichrome

b) Power supply

Battery eliminator

Regulated output: 1.2, 2, 4, 6, 8, 10 & 12 V

c) Ammeter

Range: 0 - 2 A

Resolution: 0.02 A

d) Voltmeter

Range: 0 - 10 V

Resolution: 0.1 V

e) Rheostat

Tube length: 300 mm

Contact: Spring loaded copper blades

Resistance wire: Nichrome

Terminals: 3 (X-0-Y)

Max. resistance: 100 Ω

f) Patch cords


Set of standard required patch cords




(c)



(d)



(f)



(e)



KAMALJEETH INSTRUMENTS

ESTD. 1990

Address: No. 610, 5th main, 8th cross Tatanagar, Bangalore - 560092, INDIA

Website: www.kamaljeeth.net, Email: labexperiments@kamaljeeth.net

3 years manufacturing warranty